

# AURELIA® Bold MAX®

Examination Gloves



<b>9789A7</b> (M)	ø 50	10 x 50
<b>9789A8</b> (L)	ø 50	10 x 50
<b>9789A9</b> (XL)	ø 50	10 x 50
<b>9789A0</b> (XXL)	ø 50	10 x 50



**Nitrile**



**Powder Free**



**Black**



**7.5 +/-0.2g**



**Diamond-Textured**




**5 Years**

Premium quality black nitrile examination gloves with an excellent diamond grip finish offering a greater tensile strength and increased protection than both Latex and Vinyl

- 6 Mil thickness
- 7.5g in weight
- Diamond Textured
- Latex Free
- Ambidextrous
- AQL 1.5

## Sizes & Dimensions



M	95 ± 10 mm	min. 240 mm
L	110 ± 10 mm	min. 240 mm
XL	120 ± 10 mm	min. 240 mm
XXL	≥ 120 mm	min. 240 mm

## Single wall thickness



- ① min. 0.20 mm
- ② min. 0.16 mm
- ③ min. 0.08 mm

## Physical properties

	Unaged	Aged
Tensile Strength	min. 14.0 MPa	min. 14.0 MPa
Elongation	min. 500 %	min. 400 %
Strength at Break	min. 6N	min. 6N

Aurelia Gloves conforms and complies with:

- The general safety and performance requirements of FDA Medical Device Regulation for Class 1 medical devices (21 CFR 880.6250) and complies with all general controls (section 513(a)(1)(A) of the Federal Food, Drug, and Cosmetic Act (Act); 21 U.S.C. 360c(a)(1)(A))
- The standard specification requirements for ASTM D6319- Freedom From Holes, Physical Dimensions Test, Physical Requirement Test and Packaging
- The standard specification requirements for ASTM D6978-05 - Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
- EEC regulations concerning the conformity of materials and products that are allowed to come into contact with food. In accordance with Regulation EEC 1935/2004, Regulation EC 10/2011 & Regulation (EC) No 2023/2006.
- The gloves are certified manufactured according to ISO 9001:2015 and ISO 13485:2016 Quality Management Systems